

Renewable Energy Lab Weather Station

Initial Design Report Template

Chenxi Dong: CAD Engineer

Rowan McCullough: Test Engineer and Financial Manager

Ian Torp: Project Manger

Shutong Wang: Manufacturing Engineer

Summer 2025-Fall 2025



Project Sponsor / Instructor: Carson Pete

DISCLAIMER

This report was prepared by students as part of a university course requirement. While considerable effort has been put into the project, it is not the work of licensed engineers and has not undergone the extensive verification that is common in the profession. The information, data, conclusions, and content of this report should not be relied on or utilized without thorough, independent testing and verification. University faculty members may have been associated with this project as advisors, sponsors, or course instructors, but as such they are not responsible for the accuracy of results or conclusions.

EXECUTIVE SUMMARY

The Renewable Energy Lab at Northern Arizona University has requested that we design a weather station for academic use. Professor Carson Pete approached us with the proposal and has provided us with his requirements. The weather station will potentially collect six different types of weather data and allow this to be accessed remotely at any time. The six different weather data types include rainfall, air temperature, barometric pressure, humidity, solar irradiance, and wind speed. The only weather data that our client may have us not collect from these six is rainfall data. As the project continues, we will decide with the client whether or not to collect rainfall data.

The weather station will feed all the sensors' outputs into a Raspberry Pi located inside the Renewable Energy Lab. This Raspberry Pi was provided and is a Raspberry Pi 3+. It will collect all the data from the sensors and then convert the raw data into readable data. Each sensor will need its own code within the Pi to obtain the readable data, as each sensor is calibrated differently. Some may require a table to interpret the data while others may only require a formula. Once the data is readable, we decided that it will create 10-minute averages of each sensors' data that will later be converted into 60-minute averages that get uploaded into the database and website. This is to avoid having the Raspberry Pi store a massive amount of data, while also providing accurate readings. Each 10-minute average will also be uploaded to the website, replacing the previous average so that users can access real-time weather data.

As for structural design, the lab has provided us with a tall metal truss which we may mount our sensors upon. The platform atop this truss will utilize a diamond shaped crossbar with diagonal supports mounted to the top of this platform. Every sensor except for the anemometer will be mounted at this height, as it would reach their respective standards. The anemometer will be attached via boom arm to a separate existing tower. This design does not currently support having a rain gauge, but it would not be difficult to install if the client requests it. Each sensor on the top of the truss will be mounted at the vertices of the diamond crossbar so as to maintain three points of contact to ensure stable mounting.

At the time of writing this document, the project is in the design phase. There has been no construction or prototype created yet. Each item of interest will be thoroughly explained in the following paper. Our current budget includes approximately \$3000 of existing equipment, \$500 provided by our client for extra costs, and an expected \$300 from fundraising. The current foreseeable costs would be for the creation of the metal crossbar for the sensor mounting and a new SD card for the Raspberry Pi.

TABLE OF CONTENTS

Contents

DISCLAIMER	2
EXECUTIVE SUMMARY	3
TABLE OF CONTENTS	4
1 BACKGROUND	5
1.1 Project Description	5
1.2 Deliverables	6
1.3 Success Metrics	6
2 REQUIREMENTS	7
2.1 Customer Requirements (CRs)	7
2.2 Engineering Requirements (ERs)	7
2.3 House of Quality (QFD)	8
3 Research Within Your Design Space	9
3.1 Benchmarking	9
3.2 Literature Review	9
3.2.1	9
3.2.2	11
3.2.3	13
3.2.4	14
3.3 Mathematical Modeling	15
3.3.1	15
3.3.2	15
3.3.3	16
3.3.4	17
4 Design Concepts	18
4.1 Functional Decomposition	18
4.2 Concept Generation	20
4.3 Selection Criteria	25
4.4 Concept Selection	26
5 CONCLUSIONS	28
6 REFERENCES	29
7 APPENDICES	31
7.1 Appendix A: Pseudocode	31

1 BACKGROUND

This chapter of the report will discuss the initial importance of the project. The first section will describe the project proposal and explain why the project is important. It will also discuss the budget and

fundraising targets of the project. The next section will discuss the major deliverables for the course, covering each deadline for the project. The final section of this chapter will discuss our success metrics for the project. This will include our goals as well as the major design requirements of our project in general.

1.1 Project Description

We have been tasked with assembling a weather station for the Renewable Energy Lab (RE Lab) at Northern Arizona University (NAU). This station is required to provide current, accurate weather data readings for academic and scientific use. While the RE Lab used to have functioning weather data collectors, they no longer function. Additionally, the previous collection required the professor to go into the RE lab itself and read the current data. There was no way to access this data remotely, nor was there a way to access previous data. Our aim is to fix this problem. Our project will collect data to display and store online for multiple years. This will allow us to alleviate the headache of collecting weather data from the RE Lab. While our weather station will not provide a forecast for upcoming weather events, the data collected will be sufficient to make such determinations. Weather stations are important installations, typically for the purpose of predicting future weather events, disaster preparedness, and agricultural information. Our weather station will collect data on temperature, rainfall, humidity, air pressure, solar radiation, and wind speed. The classes that our project should be the most influential for are the Renewable Energy course at NAU and the Wind Power course. The students and faculty will be able to access our data to perform calculations and other deliberations.

Our budget for this project includes around \$3000 of preexisting equipment, an additional \$500 dollars from our client for potential additional costs, and an expected \$300 in fundraising from the team. The team already has access to the equipment and has not run into any unexpected costs yet. One charge that may occur is a replacement SD card for the Raspberry Pi. The current SD card functions well, but we may need to replace it for the purpose of accessing the Pi. This cost would range anywhere from \$6 to \$50. With the fundraising target being so low, the team is prepared to provide the funds themselves if they are unsuccessful in finding external sponsors.

1.2 Deliverables

The project aims to deliver a functional modular weather station to meet the needs of the Renewable Energy Laboratory near NAU. The weather station will be designed to measure six key environmental variables: temperature, humidity, wind speed, air pressure, solar irradiance, and rainfall. Final deliverables include:

- A fully assembled solar-powered weather station utilizing a Raspberry Pi for data processing and wireless transmission.
- Robust sensor integration through modular mounting and standardized interfaces, following WMO siting guidelines.

- A cloud-based data display dashboard built with Adafruit IO capable of displaying real-time and historical measurements.
- Sensor calibration models and at least one working mathematical modeling example (e.g., humidity-to-voltage conversion or wind speed calculation).
- Custom data management code written in Python to read, process, average, and upload sensor values.
- A complete CAD design of the sensor support structure, paying attention to maintenance, sensor replacement, and weatherization.
- A detailed technical report outlining the design process, benchmark comparisons, concept generation, modeling, and test results.

1.3 Success Metrics

The success of this project will be evaluated based on functional performance and how well it meets the engineering goals described throughout the report. The following criteria will be used:

- **Measurement Accuracy:** All sensor outputs should meet the manufacturer's specified accuracy standards under test conditions.
- **Data Uptime:** The weather station should demonstrate consistent data collection and transmission, with at least 90% successful uploads within a 24-hour test window.
- **Modularity and Maintenance:** Sensors must be individually replaceable and independently verifiable in the event of a failure.
- **Online Visualization:** All six variables must be clearly displayed on the Adafruit IO dashboard with appropriate units and labels.
- **Structural Stability:** The tower and sensor mounts must be able to withstand moderate outdoor wind, rain, and temperature variations typical of Flagstaff.
- **Documentation Quality:** All code, CAD models, and calibration methods must be clearly organized and submitted in the final report format.

2 REQUIREMENTS

This section will outline the customer needs and engineering requirements of the weather station project. The customer needs were provided by our client with levels of importance assigned to each. The engineering requirements were identified as a team and are quantifiable. These were used to create the QFD which compares and identifies the relationships between each other and helps to identify what our main focuses should be. It was also used to grade our benchmarks on how well they would accomplish the needs of our project.

2.1 Customer Requirements (CRs)

Measurement of Key Weather Parameters - Station will measure temperature, humidity, wind speed and direction, barometric pressure, and solar irradiance. Will potentially measure rainfall as well.

Data Transmission - Data collected will be transmitted via internet.

Remote Data Access - Live and stored data should be accessible through a web interface.

Renewable Power Supply - Any components which require power must run on solar energy.

Weather Durability - Station must withstand outdoor weather conditions.

Low Maintenance - Station should require less than 2 hours of maintenance per year.

Sensor Accuracy - Sensor readings should be properly calibrated.

User Friendly - User interface should be easily navigable.

Ease of Installation - Installation should require minimal tools or training.

Low Cost - Station should be cost effective and within budget.

Safety Compliance - Must comply with relevant electrical and operational safety standards.

Data Storage - Data should be stored in an accessible and organized database for at least one year.

2.2 Engineering Requirements (ERs)

Long Term Data Storage - Database should log data in an organized manner over the course of 4 years.

Increased Data Accuracy - Sensor readings should be highly accurate, within 3% or less. Data average calculations should be properly computed.

Multiple Wind Speed Readings - Wind speed and direction should be measured at both standard height and atop existing tower at lab providing at least 2 readings.

Measured at Industry Standards - Sensors should be properly positioned according to industry standards.

Proper Calibration - Sensors should be properly calibrated to upload accurate data from raw readings within 3% of the true values.

Measurement of All Data Types - Station should record 5 data types including temperature, pressure, humidity, solar irradiation, and wind data.

Low Power Requirement - Station should be capable of fully operating under existing solar generated power means located at lab, with a target of 0.2 kWh per day or less.

2.3 House of Quality (QFD)

System QFD			Project: RE Weather Station QFD						
			Date: 6/16/2025						
1	Long Term Data Storage		<div>Legend</div> <div>A KFLG https://forecast.weather.gov/MapClick.aspx?lat=39&lon=-98</div> <div>B Tempest https://www.tempestinc.com/</div> <div>C Davis Vantage https://www.davisinc.com/</div>						
2	Increased Data Accuracy								
3	Multiple Windspeed Readings								
4	Measured at Industry Standards								
5	Proper Calibration								
6	Temp, Pressure, Humidity, Wind spd/dir, Solar Irradiance								
7	Low Power Requirement								
			</						

weather station provides readings for temperature, humidity, wind speed, barometer, dew point, visibility, and heat index. It also displays the latitude and longitude of the weather station, as well as its elevation. It also clearly displays when the last update occurred for the website. These updates occur every 30 minutes. A three-day and seven-day history are both available, which we would like to emulate on our own website. The website itself is clear and user-friendly, which are important design features for our team. It uses both line graphs and tables to display all the collected weather data, which is something we plan to do as well. This benchmark will act as an outline for our website as well as a data confirmation system for our own readings. If our readings are significantly different from the readings made by this station, we will need to reevaluate our approach.

Our second benchmark is the Tempest Home Weather Station, a consumer product available online [12]. This single unit device costs about \$350 and claims to measure many of the weather parameters we wish to measure. The product also includes a cell phone app which allows you to access the data from the device as well as a 10-day forecast. While the device seems incredibly versatile for its cost and portability, the station does not appear to be equipped with many of the sensors needed to make the readings it claims to. For example, there is no anemometer or wind vane of any kind attached to the unit, yet it claims to measure wind speed and direction. While the app is user friendly and provides a lot of information, there is no database feature to record past weather patterns. Additionally, some reviews of the station seem to believe that many of the ‘readings’ it claims to make are just pulled from another source and posted as coming from the personal device. Regardless of the validity of the readings, this cost-effective personal product is something we must strive to beat to justify our weather station.

Our third benchmark chose Davis Vantage Pro2 as a reference. This is a high-end weather station widely used in agriculture, scientific research and universities. Its main features include: integrated temperature, humidity, wind speed and other sensors; support wireless data transmission up to 300 meters; solar and battery-powered, suitable for field deployment. Its structure is tower-shaped and the anemometer is installed on the top to reduce interference. The modular design can also support additional functions such as ultraviolet rays. Its performance is very consistent with the goals of our project.

3.2 Literature Review

3.2.1 Ian Torp

[1] Setra Systems, Inc, “Barometric Pressure Sensors | Setra Systems,” *Setra.com*, 2025. <https://www.setra.com/product/pressure/barometric> (accessed Jun. 18, 2025). This source is a storefront for barometric pressure sensors. For the purpose of our project, this has been used to help determine some benchmarks for the pressure sensor specifically. It displays 5 different models and compares them to one another using the criteria of pressure fittings, electrical terminations, output, media compatibility, thermal effect, compensated temperature range, operating temperature, accuracy FS, ranges in PSI, sensing technology, and sample applications. This will become more useful to us as we begin to tinker with our provided barometric pressure sensor. This website also explained some of the basics of barometric pressure sensors, which was useful in determining exactly why this sensor is important.

[2] N. US Department of Commerce, “Standards and Policy,” *www.weather.gov*. <https://www.weather.gov/coop/standards>. This is a government-run website which has a number of useful resources. The one used for this project was NWSI 10-1302: Requirements and Standards for NWS Climate Observations. This document discusses the site and exposure standards of weather stations, as well as more specific air temperature measurement standards and precipitation measurement standards. It

also discusses soil temperature measurement standards and pan evaporation measurement standards, although that is not useful information for this project. This is where we learned that thermometers must be able to measure between –20 degrees Fahrenheit and 115 degrees Fahrenheit, and that precipitation measurements must be able to reach 20 inches. The discussion of each sensors' standards has informed our decision making throughout our design process.

[3] *World Meteorological Organization, Guide to meteorological instruments and methods of observation., Volume 1. Geneva, Switzerland: World Meteorological Organization, 2008.* This is my most important source so far. Each chapter in this book explains a different measurement and its standards. Each chapter discusses different sensors and tools to measure each weather feature. It goes in-depth on everything from standards to calibration. I have already utilized chapters one through seven. These are, in order: General, Measurement of Temperature, Measurement of Atmospheric Pressure, Measurement of Humidity, Measurement of Surface Wind, Measurement of Precipitation, and Measurement of Radiation. This is a vital source that will continue to contribute to our understanding and development of the weather station.

[4] ISO-CAL, "What is a Pyranometer? 10 Important Points to Consider.," *isocalnorthamerica.com*, Jan. 16, 2023. <https://isocalnorthamerica.com/what-is-a-pyranometer/> This website was used to better understand pyranometers. It discussed what exactly solar irradiance is and what it is used for, which are things like climatology and solar power generation. It described that they measure both diffuse and direct sunlight to record the amount of solar energy reaching a surface per unit area. Typically, this measurement is in $\frac{W}{m^2}$. It also briefly mentions the two common types of pyranometers as well as the importance of calibration. It does not discuss directly how to calibrate, as they are trying to sell their calibration services. This source has allowed us to understand pyranometers in simpler terms than discussed in the book mentioned above.

[5] C. Gittins, "Considering the energy consumption of a Raspberry Pi," *IOT Insider*, Sep. 23, 2024. <https://www.iotinsider.com/news/considering-the-energy-consumption-of-a-raspberry-pi/> This website was used to help determine how much power consumption our weather station will have overall. This was used in my first mathematical modelling exercise, which will be discussed in section 3.3 of this report. This website said that a Raspberry Pi will typically use 5W with normal load.

[6] NiuBol, "How much power does a weather station use ? ," *Niubol.com*, 2024. <https://www.niubol.com/Product-knowledge/How-much-power-does-a-weather-station-use.html>. This website was also used to complete my mathematical modeling calculations. Here, they talked about how much power weather measuring sensors use typically. It was very low, as each sensor is expected to only use .3W each hour. NiuBol is itself a company that sells weather stations, but it was a useful source for calculation and comparison nonetheless.

[7] US, "National Weather Service," *Weather.gov*, 2025. <https://forecast.weather.gov/MapClick.php?lat=35.198140000000002&lon=-111.65112499999998> (accessed Jun. 18, 2025). I used this source as my benchmark weather station. The weather station I am choosing to be the benchmark is the Flagstaff Pulliam Airport weather station. This weather station will act as a benchmark for both our website design and each sensors' accuracy. We will compare our readings to theirs, in the hopes of the readings being similar. This benchmark was already discussed above in section 3.1.

[8] J. Portilla, “Python Bootcamps: Learn Python Programming and Code Training,” Udemy, 2019. <https://nau.udemy.com/course/complete-python-bootcamp> (accessed Jun. 22, 2025). I used this Udemy course to teach me Python. I have no coding experience in Python and needed to start from the beginning. I have continued to use this source as a means to learn how to code and will continue to do so. This project will require a great amount of coding, so this is a vital resource for me.

[9] Raspberry Pi, “Build Your Own Weather Station,” *Raspberrypi.org*, 2017. <https://projects.raspberrypi.org/en/projects/build-your-own-weather-station/0> (accessed Jun. 27, 2025). The Raspberry Pi website has been very helpful in the design aspect of this project so far. This website discusses how to make a weather station out of a specific kit that Raspberry Pi sells. We will be using this as an outline for our own coding in the project, as it discusses the entire set up of a weather station from start to finish. We will not be copying it, as we do not have the kit that was used.

[10] Raspberry Pi, “Raspberry Pi Documentation - Getting Started,” *www.raspberrypi.com*. <https://www.raspberrypi.com/documentation/computers/getting-started.html> (accessed Jun. 27, 2025). I used this website to help me understand how to access and use a Raspberry Pi. When I tried to access the provided device, it unfortunately had a password on it. Luckily, this website discusses a number of workarounds, including a full reset or just replacing the SD card within it.

[11] Raspberry Pi, “Raspberry Pi OS,” *Raspberry Pi*, 2025. <https://www.raspberrypi.com/software/> (accessed Jun. 27, 2025). To actually activate the Raspberry Pi after replacing or wiping the SD card, we would need to upload the Raspberry Pi operating system. This website discusses how to do so as well as providing the operating system download and documentation. This will be used to fully access the Pi.

3.2.2 Rowan McCullough

[3] World Meteorological Organization, *Guide to meteorological instruments and methods of observation., Volume 1*. Geneva, Switzerland: World Meteorological Organization, 2008. This resource has been shared between several members of our group as it contains a ton of vital information for our project. I primarily utilized this resource to find and understand the industry standards for weather measurement across all types of sensors we will be utilizing. Additionally, it outlines the various subtypes of each of these sensors and their strengths and weaknesses. This source provided me with the equation for a specific thermometer output calculation shown below in section 3.3.

[12] “Tempest Weather Station,” Tempest, 2024. https://shop.tempest.earth/products/tempest-msclkid=1236cca887f915b3bbd3a4f8024b2f08&utm_source=bing&utm_medium=cpc&utm_campaign=Bing-Search-B2C-NB-Broad-US&utm_term=home%20weather%20station&utm_content=Weather%20Station (accessed Jun. 18, 2025). This source is the purchase page for the Tempest Home Weather Station; one of our benchmarks mentioned previously. It was used to understand the features and components of the weather station. It also includes customer reviews which were helpful in validating the product's performance.

[13] A. Overton, “A Guide to the Siting, Exposure and Calibration of Automatic Weather Stations for Synoptic and Climatological Observations,” 2009. Accessed: May 01, 2024. [Online]. Available: <https://www.rmets.org/sites/default/files/2019-02/aws-guide.pdf> This source is a published paper from a member of the Royal Meteorological Society which outlines the calibration and long-term considerations of various weather sensors. It served as a cross-reference to other calibration research as

well as a guide to the potential obstacles to maintaining accurate readings over a long period of time. It also contains a lot of information on data archiving that may be useful to us down the line.

[14] Wisconsin DNR, “Calibration & barometric pressure | Wisconsin DNR,” *dnr.wisconsin.gov*. <https://dnr.wisconsin.gov/topic/labCert/BODCalibration2.html> This website from the Wisconsin Department of Natural Resources discusses barometric pressure and the variations in readings based on environmental and elevation factors. Since most pressure data will be corrected to altitude, it is important to understand this process when comparing our readings to nearby stations and when calibrating our database.

[15] R. Coquilla, J. Obermeier, and B. White, “American Wind Energy Association,” 2007. Accessed: May 16, 2023. [Online]. Available: <https://research.engineering.ucdavis.edu/wind/wp-content/uploads/sites/17/2014/03/AWEA-2007-Final-Paper.pdf> This research paper was published by graduate students from University of California, Davis. It highlights uncertainty factors when calibrating an anemometer and the various roots of uncertainty that may arise and how to calculate them. It also contains test data from an anemometer calculation performed in a wind tunnel. This was an important resource for our understanding of anemometer calibration and a guideline for my uncertainty calculation self-learning assignment.

[16] “Login - CAS – Central Authentication Service,” *Udemy.com*, 2025. Available: <https://nau.udemy.com/course/statistics-intro/learn/lecture/35671972#overview>. [Accessed: Jun. 25, 2025] This Udemy course was my primary self-learning resource when refreshing myself of statistics and how uncertainty arises and can be calculated. Not all sections were utilized, just sections three, four, eight, and ten. These gave me an important baseline understanding of statistical processes to help me build a MATLAB program to calculate uncertainty in an anemometer reading.

[17] EngineerItProgram, “Experimental Uncertainty,” *YouTube*, Mar. 18, 2013. Available: https://www.youtube.com/watch?v=xJBta_HWTRc. [Accessed: Jun. 26, 2025] This YouTube video paired with the Udemy course gives me more specific information on how engineers calculate and handle uncertainty in their designs and experiments. This contributed to my self-learning assignment and will help us handle the inevitable uncertainty that will arise in our data down the line.

3.2.3 Shutong Wang

[18] National Instruments, “NI DAQ Sensor Calibration Guide,” *National Instruments*, 2021. Accessed: May 10, 2024. [Online]. Available: <https://www.ni.com/en-us/innovations/sensor-calibration.html>

This technical guide outlines procedures and best practices for calibrating sensors in data acquisition systems, especially using microcontrollers like the NI USB DAQ or Raspberry Pi. It helped us understand calibration drift and recommended frequency of recalibration.

[19] Adafruit Industries, “How to Connect Weather Sensors to Raspberry Pi,” *Adafruit Learning System*, 2022. Accessed: Apr. 25, 2024. [Online]. Available: <https://learn.adafruit.com/pi-weather-station/overview>

This online tutorial provides hardware wiring diagrams and Python libraries for interfacing weather

sensors (e.g., DHT22, BMP180) with Raspberry Pi boards. It guided us through the sensor setup and initial data acquisition stages.

[20] M. A. Islam and R. Haque, “IoT-Based Real-Time Weather Monitoring System Using Raspberry Pi,” in *Proc. Int. Conf. on IoT and Applications*, 2020, pp. 44–49.

This conference paper presents a complete prototype for an IoT weather station using a Raspberry Pi and discusses methods for storing and visualizing long-term weather data. It helped validate our design choices and inspired our data storage approach.

[21] S. Pandey et al., “Analysis of Weather Monitoring Systems Using Cloud Integration,” *Journal of Sensor Networks and Data Communication*, vol. 10, no. 2, pp. 100–105, 2021.

This paper focuses on integrating weather stations with cloud platforms (like AWS and ThingSpeak). It supported our decision to use online dashboards for data visualization and remote monitoring.

[22] Ambient Weather, “Installation Guide for Ambient Weather WS-5000,” *Ambient Weather*, 2023. Accessed: May 15, 2024. [Online]. Available: <https://ambientweather.com/ws5000manual>

This commercial user manual describes sensor placement, calibration, and protective enclosure requirements for outdoor weather stations. It informed our decisions on shielding and mounting components against wind and precipitation.

[23] Texas Instruments, “*Thermal Design Considerations for Enclosures*,” Application Report SPRA953, 2009. Accessed: July 14, 2025. [Online]. Available: <https://www.ti.com/lit/an/spra953/spra953.pdf>

This technical application note discusses how enclosure material, surface color, and exposure to solar radiation can significantly influence internal temperatures in outdoor devices. It emphasizes the importance of minimizing heat absorption and suggests design strategies such as reflective coatings and sunshades. This resource guided our material and surface treatment decisions for the weather station housing to reduce passive solar heating.

[24] IPC, “*Standard for Determining Current-Carrying Capacity in Printed Board Design (IPC-2152)*,” IPC, 2009. Accessed: July 14, 2025. [Online]. Available: <https://www.ipc.org/TOC/IPCT-2152.pdf>

This industry-standard document outlines methods to estimate internal temperature rise of electronic systems based on their power consumption and thermal environment. It highlights how enclosure design and airflow affect component longevity and performance. We used this reference to assess internal heating risks in our weather station and to support decisions regarding ventilation and spatial layout of sensors and the processor board.

3.2.3 Chenxi Dong

[25] “Printed humidity sensors,” Encyclopedia.pub. <https://encyclopedia.pub/entry/7493>. This entry provides an overview of the three main types of humidity sensors: capacitive, resistive, and printed. It is particularly useful for distinguishing between their operating principles and basic applications. This resource helped me understand which sensors are best for long-term environmental monitoring. For example, capacitive sensors are known for their balance of accuracy, durability, and response time, making them a prime candidate for our station design.

[26] “Review of Printed Humidity Sensors,” MDPI Sensors. <https://www.mdpi.com/2079-4991/13/6/1110>

This scientific paper provides a detailed technical comparison of printed humidity sensors, including their fabrication methods and sensing mechanisms. While we do not intend to use printed sensors in our designs, this paper provides an important benchmark for understanding cost-benefit trade-offs and short-term use advantages. This is particularly useful in concept evaluation and comparison with resistive and capacitive types.

[27] “Capacitive vs Resistive Humidity Sensors,” Encyclopedia.pub. – <https://encyclopedia.pub/entry/7493>

This section specifically focuses on the comparison of capacitive and resistive sensors. It explains in detail how capacitive sensors work by detecting changes in dielectric constant, while resistive sensors work by tracking changes in the resistance of hygroscopic materials. I used this source when calculating relative humidity using a capacitance-based formula that assumes the sensor follows linear dielectric behavior.

[28] BOMS Review – Comparison of sensing layers – <https://www.ias.ac.in/article/fulltext/boms/045/0238>

This resource discusses the materials used in humidity sensor construction and how they affect response time and accuracy. It is very useful when considering calibration strategies, as drift of these materials over time can affect long-term performance. It also validates our decision to consider sensor replacement cycles in our final deployment plan.

[29] “Humidity and Dew Point,” National Weather Service. https://www.weather.gov/media/epz/wxcalc/dewpoint_rh.htm

This government-hosted calculator tool is used to model relative humidity and dew point based on temperature and RH data. I referenced it to validate my sensor output calculations. It also provides insight into how the sensor interprets the environmental data into meaningful weather measurements.

[30] Ma, H. et al., “Graphene-Based Humidity Sensors,” arXiv. <https://arxiv.org/abs/2410.02255>
This paper introduces a new class of printed humidity sensors using graphene. While still in the research phase, it demonstrates future directions for ultra-sensitive and flexible sensors, which could potentially enhance smart wearables or high-resolution atmospheric monitoring. I use this to contextualize the current sensor options we are evaluating with options that may become viable in future iterations of our space station.

[31] NOAA Sensor Siting Standards – <https://www.weather.gov/coop/standards>
The official guide outlines the proper installation protocol for humidity sensors, including height (1.25-2 m above the ground), shielding from direct sunlight, and location relative to vegetation.

[32] WMO Guide to Meteorological Instruments – https://library.wmo.int/index.php?lvl=notice_display&id=12407

This is where I learned about long-term stability, calibration requirements, and acceptable error ranges for humidity sensors. I specifically used its advice to ensure accurate RH data in my automated weather system.

3.3 Mathematical Modeling

3.3.1 Power Requirements – Ian Torp

It was important to estimate the power consumption of the weather station. To do so, we examined what the typical power consumption of both a Raspberry Pi and weather sensors would be under normal load. A Raspberry Pi typically consumes 5W under normal load [5], whereas each sensor will typically only consume .3W per hour [6]. To figure out how much power would be consumed daily, we need to calculate how many kWh are consumed. There are 24 hours in each day, so we take that and multiply it by both the .3W per sensor and the 5W for the Raspberry Pi. Since we have 7 total sensors potentially, the equation becomes:

$$24\text{hrs} \cdot 7(.3\text{W}) + 24\text{hrs} \cdot 5\text{W} \quad (1)$$

This means the weather station consuming .1704 kWh per day. This becomes 62.192 kWh per year. This is very low power consumption. A simple, small 60 W solar panel is more than enough to support this weather station entirely every day with only 5 hours of total sunlight. A 60 W solar panel will produce about 300Wh per day with only 5 hours of sunlight, which is more than enough to support the station.

3.3.2 Thermometer Output Calculation – Rowan McCullough

Our station is expected to use an Electrical Thermometer (PRT) which measures temperature through the resistance of pure metal. The resistance of pure metals has a nearly linear relationship with temperature across the meteorological range [3]. Different electrical thermometers are categorized based on the metals they use, with the most common being the Pt100. The Pt100 sensor also has sub-categories which are categorized by standard tolerances such as IEC 60751 or ASTM E1137 [3] [13]. These standards define the resistance and sensitivity curves which are used to convert outputs into temperature values. The electrical resistance of a PRT can be represented by the Callender-Van Dusen Equation shown below.

$$R = R_0 \left(1 + At + Bt^2 + C(t - 100)t^3 \right) \quad (2)$$

In this equation, R_0 represents the electrical resistance at 0 °C and t represents the temperature in degrees Celsius. The coefficients A , B , and C depend on the standard specification. Assuming the IEC 60751 specification, $R_0=100\Omega$, $A=3.908 \times 10^{-3} \text{ } ^\circ\text{C}$, $B=-5.80 \times 10^{-7} \text{ } ^\circ\text{C}$ and $C=4.27 \times 10^{-12} \text{ } ^\circ\text{C}$ when temperatures are below zero and $C=0 \text{ } ^\circ\text{C}$ when above zero. In temperatures above -40 °C, the C term is negligible and thus the inverse equation for temperature can be simplified to:

$$t = -\frac{A}{2B} + \frac{A}{2B} \sqrt{1 - \frac{4B}{A^2} \left(1 - \frac{R}{R_0} \right)} \quad (3)$$

Therefore, given an output reading of 110Ω, our temperature reading would be approximately 25.34 °C.

3.3.3 Thermal Management Considerations – Shutong Wang

Outdoor electronic devices are prone to internal temperature increases due to direct sunlight, temperature fluctuations, and insufficient airflow. Excessive heat can affect sensor accuracy and accelerate component aging.

Joule heating

The internal temperature rise caused by total power loss can be estimated using the following thermal resistance model:

$$\Delta T = P * R_{\theta JA}$$

(4)

- ΔT : Device temperature rise ($^{\circ}\text{C}$)
- P : Total power consumption (W)
- $R_{\theta JA}$: Device or system junction-to-ambient thermal resistance ($^{\circ}\text{C}/\text{W}$)

Result:

If the total power consumption of the sensor is 7.1W and the thermal resistance is $10^{\circ}\text{C}/\text{W}$, the temperature rise will be approximately 71°C . If the ambient temperature is 35°C , the internal temperature of the housing may reach as high as 106°C .

Solar Radiation Heating

In outdoor applications, solar radiation can significantly increase the temperature of exposed enclosures. The absorbed power can be calculated as:

$$Q_{solar} = G \cdot A \cdot \alpha$$

(5)

- G : Solar radiation intensity (typical: $800\text{--}1000 \text{ W}/\text{m}^2$)
- A : Light-receiving area (m^2)
- α : Material absorption coefficient (black: almost 0.9, white almost 0.3)

Result:

A 0.1m^2 black box receives:

$$Q = 1000 * 0.1 * 0.9 = 90\text{W}$$

Reflective coating or sunshade design is required

3.3.4 Humidity sensor output calculation – Chenxi Dong

For capacitive humidity sensors, relative humidity (%RH) is typically determined by the change in capacitance:

$$RH = \frac{C - C_{dry}}{C_{wet} - C_{dry}} \times 100$$

(6)

Where:

C is the measured capacitance

C_{dry} = capacitance at 0% RH

C_wet = capacitance at 100% RH

Example Sensor:

C_dry = 180 pF, C_wet = 250 pF, Measured C = 216pF


$$RH = \frac{216-180}{250-180} \times 100 = \frac{36}{70} \times 100 = 51.4\% \quad (7)$$

Validation was accomplished by comparing the values to the manufacturer's calibration curve. This modeling allowed us to validate sensor sensitivity to meet the sensor tolerance of $\pm 2-3\%$ RH and guided us in performing automated calibration to compensate for drift over time.


4 Design Concepts

4.1 *Functional Decomposition*

The functional decomposition diagram is crucial for our weather station project. It breaks down the complex system into key sub-functions such as sensing, signal processing, data management, data transmission, power management, and structural protection, helping us to clarify how to implement each engineering requirement. This structured approach allows us to more efficiently assign tasks, identify technical challenges in advance (such as analog-to-digital conversion, solar power regulation, etc.), and ensure the organic integration of subsystems. In addition, the functional diagram also clearly shows the logical path from environmental input to data output, making the design easier to verify and test. The decomposition model also has good modularity, which is convenient for subsequent debugging, upgrades, or component replacement



Black Box Model




4.2 Concept Generation

The two main systems which were applicable for concept generation in our project were the structural layout of the sensors and the coding strategy used to handle the data. Each team member designed a concept for both systems. These concepts can be seen below. The concept determined to be middle of the road was chosen as the DATUM for the others to be compared against in the Pugh Chart seen in section 4.4.


4.2.1 Structural Concepts:

DATUM Design (Rowan):




This structural concept utilizes the small platform which already exists outside the RE Lab as well as the existing anemometer tower. The platform will utilize a diamond shaped crossbar with diagonal supports mounted on top of the platform. The sensors which are expected to be positioned at this height, which includes all but the anemometer, will be mounted at the vertices to provide three points of contact for stability. The anemometer will be mounted on a boom at the industry standard height of ten meters on the existing tower. Any sensors which need additional housing can be housed within the appropriate material while remaining far enough from the other sensors to avoid interference. The pros of this concept include adherence to industry standards, low cost, and durability. Some cons include potential difficulty building and installing the crossbar and boom and potential safety hazards of the crossbar having corners mounted near head height.

Design 1 (Chenxi):



The structural concept uses a modular tower structure with a height of about 3 meters, which is reinforced by tripod supports and cables. The sensors are installed in layers to facilitate independent arrangement and reduce mutual interference. The anemometer is installed at the top (3 meters) to ensure unobstructed airflow. The solar radiation meter is installed on the south side at a height of 2 meters to ensure sufficient sunlight. The barometer and thermometer are installed on the opposite side at a height of 1.5 meters. The humidity sensor is installed at the lowest level (1–1.25 meters) and is placed in a housing with ventilation and radiation shielding to avoid direct sunlight and heat interference. The data cables of all sensors are routed from the inside of the main pole, which is both wind-proof and rain-proof, and also beautiful and neat. The modular structure design facilitates replacement, calibration and transportation. Its main disadvantage is that the anemometer height does not reach 10 meters, which does not meet the industry standard for wind measurement, which may affect the accuracy of the data.


Design 2 (Shutong):



Station.

8

Design 3 (Ian):



This concept is similar to the DATUM in that it utilizes a mounted system atop the existing platform located outside the lab. The major difference lies in the mounting system used for the sensors, with this concept using a crossbar with a sensor located at the end of each rod. The pressure and humidity sensors will be mounted atop each other as they have similar exposure requirements. The major difference is the anemometer placement, with this concept having it placed on a vertical rod directly above the platform. The pros of this concept are similar to the DATUM, with ease of installation being added to the pros. The negatives of this concept include a lack of stability and durability with a single point of contact for each sensor mount and the height of the anemometer not reaching the industry standard of ten meters.

4.2.1 Pseudocode Concepts

DATUM (Ian):

The pseudocode I created was a simple outline describing the processes that would occur within the Raspberry Pi for each sensor. This is visible within Appendix A. The python-based code read the sensor data as it came in and completed the calibration calculations for each sensor accordingly. The idea was for each of these data points to be stored within the Raspberry Pi within respective sections for each sensor. I proposed that this data be uploaded to the website every 5 minutes to provide clear and up-to-date information. I also suggested that the Raspberry Pi should have external storage plugged into it so that we may have a local storage backup of all the data collected for about a year.

This design is simple and would be clear to read as an outside programmer. It would also ensure that we have backup storage for the data in the event that the website has an error occur. Unfortunately, this design would also be more expensive as it would require an external hard drive to be purchased. This design would require a lot of storage space, as 5 minutes is a much faster upload rate than the Flagstaff Pulliam Airport Weather Station's data upload rate is. This would also entail a higher power consumption than other designs may require.

Design 1 (Chenxi):

I propose a Python-based script: it reads sensor data every 10 minutes and uploads the data to the local database and server synchronously. The data of each sensor is uploaded independently, rather than uploaded together, so that when a sensor fails, it will not affect the data of other sensors. At the same time, after a day has passed, the data of the whole day will be retained in the local database for an additional day to deal with data loss. Another option is to dynamically change the data recording frequency according to environmental changes: when several data are almost the same or the difference is not big, slow down the recording frequency; when the environment changes greatly, speed up the recording frequency.

Design 2 (Shutong):

This design takes a structured approach to processing and visualizing ambient weather data using Python-based tools. Data is collected every 30 seconds from local CSV files or remote cloud sources. Each data input includes a timestamp, temperature, humidity, and air pressure readings. These values are immediately cleaned using libraries to handle missing or invalid entries and ensure consistent datetime formats. pandas

Every 30 seconds, the system adds a new record to a short-term buffer that stores up to 20 calibration readings. Every minute, it calculates a rolling average from these readings to reflect current conditions. This live value is rendered as a live chart using libraries such as or and can optionally be uploaded to a simple dashboard built with tools such as Dash or Streamlit. matplotlibseaborn

In addition to real-time updates, the system also supports hierarchical averaging of long-term summaries. Every 10 minutes, the system calculates an average from the past 10 1-minute values and stores it in a buffer. Once the six 10-minute blocks of data are collected, they are averaged again to produce a final hourly summary that can be stored for historical analysis or downloaded as a PDF report.

Visualizations include:

- A line chart shows temperature over time.
- A bar chart to compare barometric pressure values at different intervals.
- A multi-line chart showing both temperature and humidity.
- Each chart is labeled, and color styled for clarity and automatically updates to reflect the latest readings.

The system is designed to support scheduled updates and an optional dashboard interface. Users can select sensor types or time windows to view specific data sets, and the entire process can be configured to repeat automatically using CRON jobs or time-based triggers in Python.

Design 3 (Rowan):

This design will utilize simple hourly averages calculated from six ten-minute averages for database upload. The sensors will be read every 30 seconds, holding up to 20 readings which are then averaged and sent to the website as a live reading. Every time a new reading is taken, the 20th reading will be deleted and replaced, and every minute a new live reading will replace the one shown on the website. This will provide real-time, accurate readings for anyone who wants to see the current weather values

when viewing our website. After every 20 readings, a separate average will be taken and saved to a buffer which will collect them until one hour has passed. Then an hourly average is uploaded to the database for storage. This will prevent having too much data stored on the Raspberry Pi at one time, while still providing real-time and long-term data to the website. A basic overview of this process can be seen below, while the actual code structure can be seen in Appendix A.

- Every 30 sec:

Add calibrated raw data to a buffer of 20 readings.

- Every 1 min:

Compute rolling average from the last 20 readings.

Upload it to the website as the current "real-time" value.

Store it in the 10-minute buffer.

- Every 10 min:

When 10 1-minute values are collected, average them simply (equal weight).

Add this to the hourly buffer.

- Every 60 min:

When 6 ten-minute blocks are filled, average them to form the hourly summary.

Send to the database.

4.3 Selection Criteria

4.3.1 Structural Criteria

The selection criteria for the structural design were pulled directly from our engineering requirements and customer needs if they were applicable to this system. While we could not do calculations for most of them without a physical part, we discussed as a group to best determine how well each design accomplished them and how heavily they should be weighted in our decision matrix. These criteria included weather durability, safety compliance, low maintenance, easy installation, multiple wind speed readings, and adherence to industry standards.

4.3.2 Pseudocode Criteria

Similar to the structural criteria, these were pulled from our engineering requirements and customer needs if they were related to this system. While our actual code is not quantifiable for many of these criteria, we understand how well they would accomplish each of these goals and were able to determine their completion of each criterion and weigh each criterion for the decision matrix.

4.4 Concept Selection

4.4.1 Structural Concept Selection

Each design was placed into the Pugh chart, with the two top ranked designs moving on to the decision matrix.

Pugh Chart

Criteria	DATUM (Rowan)	Design 1 (Chenxi)	Design 2 (Shutong)	Design 3 (Ian)
Weather Durability	0	-1	1	-1
Safety Compliance	0	1	1	0
Low Maintenance	0	0	-1	0
Easy Installation	0	0	-1	0
Multiple Windspeed Readings	0	0	0	0
Measured at Industry Standards	0	-1	-1	-1
	Totals	-1	-1	-2
	Rank	2	2	4

The first step of the concept selection utilized was the Pugh chart, with a DATUM chosen and each design ranked as being better or worse at accomplishing each criterion than said DATUM. A 1 represents being better, while a -1 represents being worse, and a 0 represents being the same. These are then totaled to give a ranking of designs. We had design one and two tied for second, so after team discussion we decided to have design 1 move on to the decision matrix.

Criteria	Weights	Design DATUM (Rowan)		Design 1 (Chenxi)	
		Unweighted Score	Weighted Score	Unweighted Score	Weighted Score
Weather Durability	0.2	75	15	75	15
Safety Compliance	0.1	90	9	100	10
Low Maintenance	0.2	90	18	70	14
Easy Installation	0.15	50	7.5	30	4.5
Multiple Windspeed Readings	0.1	100	10	100	10
Measured At Industry Standards	0.25	80	20	60	15
Sum	1	485	79.5	435	68.5

In this decision matrix we assigned weights to each criterion, totaling to 1. We then gave scores to each design on how well they accomplished the criterion out of 100. These scores were multiplied by the weights of each criterion to give each design a score out of 100. The winning design was the DATUM design with a score of 79.5.

4.4.2 Pseudocode Concept Selection

Using the same process as the structural system, each pseudocode design was put into a Pugh chart to determine the two best designs which were then utilized in the decision matrix.

Pugh Chart

Criteria	DATUM (Ian)	Design 1 (Chenxi)	Design 2 (Shutong)	Design 3 (Rowan)
Long Term Data Storage	0	0	-1	1
Increased Data Accuracy	0	1	0	0
Remote Data Access	0	0	1	1
User Friendly	0	-1	0	0
Low Power Requirement	0	1	-1	0
Data Transmission	0	0	1	0
	Totals	1	0	2
	Rank	2	3	1

After comparing each pseudocode design against the DATUM, design 1 and design 3 were ranked the highest and advanced to the decision matrix.

Criteria	Weights	Design 1 (Chenxi)		Design 3 (Rowan)	
		Unweighted Score	Weighted Score	Unweighted Score	Weighted Score
Long Term Data Storage	0.25	75	18.75	85	21.25
Increased Data Accuracy	0.25	90	22.5	70	17.5
Remote Data Access	0.25	80	20	90	22.5
User Friendly	0.1	60	6	75	7.5
Low Power Requirement	0.1	80	8	80	8
Data Transmission	0.05	90	4.5	90	4.5
Sum	1	475	79.75	490	81.25

After weights were assigned to each criterion, scores were given to each design on their accomplishment of the criterion on a scale of zero to 100. These scores were then multiplied by the weights of each criterion to give an overall score to each design out of 100. The winning design was design 3 with a score of 81.25.

5 CONCLUSIONS

This document details the current progress on the design and development of the Renewable Energy Lab Weather Station for Northern Arizona University. The weather station has been requested to provide weather data measurements of air temperature, humidity, barometric pressure, solar irradiance, wind speed, and precipitation. The requirement for precipitation measurements is not set in stone, as the client wanted us to consider it as a possibility for now. This weather station is expected to collect accurate scientific data and have its data accessible from anywhere with an internet connection.

Professor Carson Pete has specifically requested that the data be transmitted and visible via an internet connection. He also requested that the station run off a renewable energy supply, which is a simple request as this station is located in the Renewable Energy Lab. It is also expected to be durable enough to withstand all outdoor weather conditions. Each sensor is expected to be accurate and low maintenance so that the station may be left alone while still collecting accurate data. The website the data is uploaded to is expected to be user friendly and have the capability to store the data for at least a year. We are also expected to stay within budget, as this project should be relatively low cost.

In order to adhere to these requirements, we went through a design selection process. Our final design has three main aspects. The structural aspect of the design involves the provided metal truss that is in the RE Lab as well as one of the towers stationed there. On top of the six-foot truss, a diamond shaped crossbar will be secured in a way that would provide each sensor mounted on its vertices with three points of contact. This current design does not include a precipitation measurement mechanism. The thermometer, barometric pressure sensor, pyranometer, and humidity sensor will be mounted atop this crossbar. This successfully adheres to their respective measurement standards. To meet the standards for the anemometer, it will be mounted with a boom arm 10 meters off one of the existing towers. The next aspect of this design decision is the pseudocode. The pseudocode utilizes simple hourly averages calculated from six ten-minute averages for database upload. The sensors will be read every 30 seconds, holding up to 20 readings which are then averaged and sent to the website as a live reading. Each time a new reading is taken, the oldest reading will be deleted from the Raspberry Pi. Each minute, a new live reading will be uploaded to the website, replacing the previous one. This provides real-time information that is accessible from anywhere with an internet connection. An hourly average is uploaded to the database to be stored indefinitely. This should prevent having too much data stored on the Raspberry Pi.

6 REFERENCES

- [1] Setra Systems, Inc, “Barometric Pressure Sensors | Setra Systems,” *Setra.com*, 2025. <https://www.setra.com/product/pressure/barometric> (accessed Jun. 18, 2025).
- [2] N. US Department of Commerce, “Standards and Policy,” *www.weather.gov*. <https://www.weather.gov/coop/standards>
- [3] World Meteorological Organization, *Guide to meteorological instruments and methods of observation., Volume 1*. Geneva, Switzerland: World Meteorological Organization, 2008.
- [4] ISO-CAL, “What is a Pyranometer? 10 Important Points to Consider.,” *isocalnorthamerica.com*, Jan. 16, 2023. <https://isocalnorthamerica.com/what-is-a-pyranometer/>
- [5] C. Gittins, “Considering the energy consumption of a Raspberry Pi,” *IOT Insider*, Sep. 23, 2024. <https://www.iotinsider.com/news/considering-the-energy-consumption-of-a-raspberry-pi/>
- [6] NiuBol, “How much power does a weather station use ? ,” *Niubol.com*, 2024. <https://www.niubol.com/Product-knowledge/How-much-power-does-a-weather-station-use.html>
- [7] US, “National Weather Service,” *Weather.gov*, 2025. <https://forecast.weather.gov/MapClick.php?lat=35.19814000000002&lon=-111.65112499999998> (accessed Jun. 18, 2025).
- [8] J. Portilla, “Python Bootcamps: Learn Python Programming and Code Training,” *Udemy*, 2019. <https://nau.udemy.com/course/complete-python-bootcamp> (accessed Jun. 22, 2025).
- [9] Raspberry Pi, “Build Your Own Weather Station,” *Raspberrypi.org*, 2017. <https://projects.raspberrypi.org/en/projects/build-your-own-weather-station/0> (accessed Jun. 27, 2025).
- [10] Raspberry Pi, “Raspberry Pi Documentation - Getting Started,” *www.raspberrypi.com*. <https://www.raspberrypi.com/documentation/computers/getting-started.html> (accessed Jun. 27, 2025).
- [11] Raspberry Pi, “Raspberry Pi OS,” *Raspberry Pi*, 2025. <https://www.raspberrypi.com/software/> (accessed Jun. 27, 2025).
- [12] “Tempest Weather Station,” *Tempest*, 2024. https://shop.tempest.earth/products/tempest-msclkid=1236cca887f915b3bbd3a4f8024b2f08&utm_source=bing&utm_medium=cpc&utm_campaign=Bing-Search-B2C-NB-Broad-US&utm_term=home%20weather%20station&utm_content=Weather%20Station (accessed Jun. 18, 2025).
- [13] A. Overton, “A Guide to the Siting, Exposure and Calibration of Automatic Weather Stations for Synoptic and Climatological Observations,” 2009. Accessed: May 01, 2024. [Online]. Available: <https://www.rmets.org/sites/default/files/2019-02/aws-guide.pdf>
- [14] Wisconsin DNR, “Calibration & barometric pressure | | Wisconsin DNR,” *dnr.wisconsin.gov*. <https://dnr.wisconsin.gov/topic/labCert/BODCalibration2.html>
- [15] R. Coquilla, J. Obermeier, and B. White, “American Wind Energy Association,” 2007. Accessed: May 16, 2023. [Online]. Available: <https://research.engineering.ucdavis.edu/wind/wp-content/uploads/sites/17/2014/03/AWEA-2007-Final-Paper.pdf>

- [16] “Login - CAS – Central Authentication Service,” *Udemy.com*, 2025. Available: <https://nau.udemy.com/course/statistics-intro/learn/lecture/35671972#overview>. [Accessed: Jun. 25, 2025]
- [17] *EngineerItProgram*, “Experimental Uncertainty,” *YouTube*, Mar. 18, 2013. Available: https://www.youtube.com/watch?v=xJBta_HWTRc. [Accessed: Jun. 26, 2025]
- [18] National Instruments, “NI DAQ Sensor Calibration Guide,” *National Instruments*, 2021. Accessed: May 10, 2024. [Online]. Available: <https://www.ni.com/en-us/innovations/sensor-calibration.html>
- [19] Adafruit Industries, “How to Connect Weather Sensors to Raspberry Pi,” *Adafruit Learning System*, 2022. Accessed: Apr. 25, 2024. [Online]. Available: <https://learn.adafruit.com/pi-weather-station/overview>
- [20] M. A. Islam and R. Haque, “IoT-Based Real-Time Weather Monitoring System Using Raspberry Pi,” in *Proc. Int. Conf. on IoT and Applications*, 2020, pp. 44–49.
- [21] S. Pandey et al., “Analysis of Weather Monitoring Systems Using Cloud Integration,” *Journal of Sensor Networks and Data Communication*, vol. 10, no. 2, pp. 100–105, 2021.
- [22] Ambient Weather, “Installation Guide for Ambient Weather WS-5000,” *Ambient Weather*, 2023. Accessed: May 15, 2024. [Online]. Available: <https://ambientweather.com/ws5000manual>
- [23] Texas Instruments, “*Thermal Design Considerations for Enclosures*,” Application Report SPRA953, 2009. Accessed: July 14, 2025. [Online]. Available: <https://www.ti.com/lit/an/spra953/spra953.pdf>
- [24] IPC, “*Standard for Determining Current-Carrying Capacity in Printed Board Design (IPC-2152)*,” IPC, 2009. Accessed: July 14, 2025. [Online]. Available: <https://www.ipc.org/TOC/IPCT-2152.pdf>
- [25] “Printed humidity sensors,” *Encyclopedia.pub*. <https://encyclopedia.pub/entry/7493>.
- [26] “Review of Printed Humidity Sensors,” *MDPI Sensors*. <https://www.mdpi.com/2079-4991/13/6/1110>
- [27] “Capacitive vs Resistive Humidity Sensors,” *Encyclopedia.pub*. – <https://encyclopedia.pub/entry/7493>
- [28] BOMS Review – Comparison of sensing layers – <https://www.ias.ac.in/article/fulltext/boms/045/0238>
- [29] “Humidity and Dew Point,” *National Weather Service*. https://www.weather.gov/media/epz/wxcalc/dewpoint_rh.htm
- [30] Ma, H. et al., “Graphene-Based Humidity Sensors,” *arXiv*. <https://arxiv.org/abs/2410.02255>
- [31] NOAA Sensor Siting Standards – <https://www.weather.gov/coop/standards>
- [32] WMO Guide to Meteorological Instruments – https://library.wmo.int/index.php?lvl=notice_display&id=12407

7 APPENDICES

7.1 *Appendix A: Pseudocode*

Design 1(Chenxi)-

Main Ideas:

Independent upload for each sensor.

Local backup for 24 hours.

Adaptive sampling rate based on environmental stability.

Initialization:

Load all sensor drivers.

Connect to SQLite database and server.

Define thresholds for environmental stability.

Every 10 minutes:

for sensor in sensors:

 data = read(sensor)

 timestamp = get_time()

 if sensor == "humidity":

 if abs(data - last_value) > 10 or data > 100:

 flag = "outlier"

 else:

 flag = "valid"

 save_local(sensor, data, timestamp, flag)

 else:

 save_local(sensor, data, timestamp)

 upload(sensor, data, timestamp)

Data is collected every 10 minutes.

Humidity data adds abnormal value judgment.

All data is first saved to the local database and then uploaded to the remote server.

Daily Maintenance:

```
delete_local_data(older_than=24_hours)
```

```
generate_daily_report()
```

Clean up local caches older than 24 hours every day and generate a sensor operation report for the day.

Adaptive Sampling Rate Logic:

```
if variance(sensor_data) < threshold:
```

```
    set_interval(longer) # Environment is stable → reduce sampling frequency
```

```
else:
```

```
    set_interval(shorter) # Drastic changes in the environment → Increase sampling frequency
```

Determine the sampling interval by judging the degree of data change.

Environmental stability → energy saving

Environmental fluctuations → improved resolution

Design 2(Shutong)-

Script: weather_visualization.py

Load data from CSV file

```
file_path = "weather_data.csv" data = read_csv(file_path)
```

Clean the data

Convert 'timestamp' to datetime format Remove rows with missing values

Filter temperature: Keep values between -50 and 60°C

Filter pressure: Keep values between 800 and 1100 hPa

Filter last 24 hours of data

```
now = current_datetime() start_time = now - 24 hours recent_data = data where timestamp >= start_time
```

Plot temperature chart

Create figure (10x4) Plot timestamp vs. temperature Label: "Temperature (°C)" Title: "Temperature Over Last 24 Hours" Rotate x-ticks by 45 degrees Add grid and legend Save as "temperature_last24h.png"

Plot pressure chart

Create figure (10x4) Plot timestamp vs. pressure Label: "Pressure (hPa)" Title: "Pressure Over Last 24 Hours" Rotate x-ticks by 45 degrees Add grid and legend Save as "pressure_last24h.png"

Output confirmation message

```
Print: "Charts saved: temperature_last24h.png and pressure_last24h.png"
```


Design 3 (Rowan)-

Buffers

rolling_buffer = [] # Up to 20 raw readings (~10 min at 30s intervals)

ten_minute_buffer = [] # 1 average per minute, keep 10

hourly_buffer = [] # 1 average per 10 minutes, keep 6

LOOP every 30 seconds:

data = read_and_calibrate_all_sensors()

Maintain rolling buffer

IF length of rolling_buffer == 20:

Remove oldest

Append data to rolling_buffer

Every 1 minute: upload rolling average for real-time display

IF time_since_last_minute == 60 sec:

rolling_avg = average_buffer(rolling_buffer)

upload_to_website(rolling_avg)

Append rolling_avg to ten_minute_buffer

Maintain 10-minute buffer

IF length of ten_minute_buffer == 10:

ten_min_avg = average_buffer(ten_minute_buffer)

Append ten_min_avg to hourly_buffer

Clear ten_minute_buffer

Every 60 minutes (6x10min)

IF length of hourly_buffer == 6:

hourly_avg = average_buffer(hourly_buffer)

store_hourly_avg_to_database(hourly_avg)

Clear hourly_buffer

DATUM (Ian):

IMPORTANT NOTE CODE IS HIGHLY DEPENDANT ON WHAT SENSOR IS BEING USED

Anemometer: will produce signals as it rotates. These signals will likely occur twice per rotation. With this, wind speed will be calculated as:

$Circumference = 2 * \pi * radius$

$Rotations = signals/2$

$Speed = (rotations * circumference) / time$

#this will then convert to kmph or mph through simple conversion math.

For rainfall: the signal will be produced when one bucket fills enough to tip over. Knowing the tipping point of the bucket is essential for calculating the overall rainfall.

bucket = .227mm #this is example size not real

count = 0

def bucket_tipped():

global count

Count = count+1

Print (count*bucket)

Rainfall= bucket_tipped/time

#time should measure from start of rain till end

#Reset function here

Temperature:

The thermometer can output many things depending on what it is. Assuming it outputs current it will be in mA

Reading = data input #in mA

Temp = # reading altered via reference table or specific formula.

#if using reference table, indexing and slicing very important. Additionally, the table will be uploaded to the Raspberry Pi.

Humidity

Similar to temperature, it will produce voltage or amps.

The code will take those outputs and convert them either using a reference table or a specific formula, likely provided in the instructions for the device.

Humidity = (input) + #modification based on input. Either formula or reference table

Barometric Pressure

Again, similar to temperature and entirely dependent on what sensor we have.

Pressure = (input) #modified with formula or table

Solar irradiance

#Will likely need a voltage reader, then convert that voltage to W/m^2 through calibration/ specific formula

Irradiance= (V #modified by formula) in (W/m^2)

#All of this data will be collected internally into separate tables specific to each sensor. Then, every 5 min (or sooner for certain devices) it will upload this to the website where it will then be sorted properly and displayed in a user-friendly manner.